	Application No.	Applicant(s)	
	10/767,637	MASON, MICHAEL D.	
Notice of Allowability	Examiner	Art Unit	
	Annette R. Reimers	3733	
The MAILING DATE of this communication All claims being allowable, PROSECUTION ON THE MERIT herewith (or previously mailed), a Notice of Allowance (PTOL NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATEN of the Office or upon petition by the applicant. See 37 CFR	S IS (OR REMAINS) CLOSED in L-85) or other appropriate commu NT RIGHTS. This application is s	this application. If not included nication will be mailed in due course. <b>TH</b>	
1. This communication is responsive to <u>09/28/06</u> .			
2. The allowed claim(s) is/are 1 and 2.			
3. Acknowledgment is made of a claim for foreign prior  a) All b) Some* c) None of the:  1. Certified copies of the priority documents  2. Certified copies of the priority documents  3. Copies of the certified copies of the priority International Bureau (PCT Rule 17.2(a)).  * Certified copies not received:  Applicant has THREE MONTHS FROM THE "MAILING DAN noted below. Failure to timely comply will result in ABAND THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.  4. A SUBSTITUTE OATH OR DECLARATION must be a INFORMAL PATENT APPLICATION (PTO-152) which so the including changes required by the Notice of Drafts (a) including changes required by the Notice of Drafts (b) including changes required by the attached Exampaper No./Mail Date  (b) including changes required by the attached Exampaper No./Mail Date  (b) Including changes required by the attached Exampaper No./Mail Date  (b) DEPOSIT OF and/or INFORMATION about the cattached Examiner's comment regarding REQUIREMING.	have been received.  have been received in Application by documents have been received.  ATE" of this communication to file ONMENT of this application.  submitted. Note the attached EXA in gives reason(s) why the oath or must be submitted.  sperson's Patent Drawing Review inner's Amendment / Comment or CFR 1.84(c)) should be written on the in the header according to 37 CFR deposit of BIOLOGICAL MATE	in No  in this national stage application from the requirements a reply complying with the requirements.  MINER'S AMENDMENT or NOTICE OF declaration is deficient.  ( PTO-948) attached  in the Office action of the drawings in the front (not the back) of R 1.121(d).  RIAL must be submitted. Note the	
Attachment(s)  1. Notice of References Cited (PTO-892)  2. Notice of Draftperson's Patent Drawing Review (PTO-9)  3. Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date	948) 6. ☐ Interview Su Paper No./I 7. ☐ Examiner's A	ormal Patent Application mmary (PTO-413), Mail Date Amendment/Comment	
<ol> <li>Examiner's Comment Regarding Requirement for Deposition of Biological Material</li> </ol>	osit 8. 🛭 Examiner's 9. 🗌 Other	Statement of Reasons for Allowance	

EDUARDO O. ROBERT SUPERVISORY PATENT EXAMINER



Application/Control Number: 10/767,637

Art Unit: 3733

## REASONS FOR ALLOWANCE

The following is an examiner's statement of reasons for allowance:

The claims in the instant application have not been rejected using prior art because no references, or reasonable combination thereof, could be found which disclose, or suggest a method for fusing a first vertebra to a second adjacent vertebra. the method comprising: a) providing an implant, the implant comprising a body having first and second opposite surfaces, wherein each of the surfaces includes at least one protruding member for securing the body to an adjacent vertebra, each protruding member of the implant characterized by a member longitudinal axis, each member having a profile perpendicular to the member's longitudinal axis that includes a generally arcuate portion that encompasses more than one hundred and eighty degrees of curvature as measured from the center of curvature, and each of the surfaces and protruding members including a bioactive coating; b) forming at least one keyway in the first vertebra corresponding to each of the at least one protruding members on the first surface, and at least one keyway in the second vertebra corresponding to each of the at least one protruding members on the second surface, wherein each keyway is characterized by a keyway longitudinal axis and has a profile perpendicular to the keyway's longitudinal axis including a generally arcuate portion that encompasses more than one hundred and eighty degrees of curvature as measured from the center of curvature; and c) inserting the implant between the first vertebra and the second vertebra in a manner so that each protruding member slides into the corresponding keyway; such that movement between the implant and the first vertebra along any

Application/Control Number: 10/767,637

Art Unit: 3733

perpendicular to the longitudinal axis of the at least one keyway in the first vertebra is

prevented, as set forth in claim 1.

Any comments considered necessary by applicant must be submitted no later

Page 3

than the payment of the issue fee and, to avoid processing delays, should preferably

accompany the issue fee. Such submissions should be clearly labeled "Comments on

Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the

examiner should be directed to Annette R. Reimers whose telephone number is (571)

272-7135. The examiner can normally be reached on Monday-Friday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's

supervisor, Eduardo Robert can be reached on (571) 272-4719. The fax phone number

for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the

Patent Application Information Retrieval (PAIR) system. Status information for

published applications may be obtained from either Private PAIR or Public PAIR.

Status information for unpublished applications is available through Private PAIR only.

For more information about the PAIR system, see http://pair-direct.uspto.gov. Should

you have questions on access to the Private PAIR system, contact the Electronic

Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a

USPTO Customer Service Representative or access to the automated information

system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

AR

EDUARDOC. ROBERT